

	Bait Al Batterjee Medical Trading Co.
	Saudi German Hospital Group
	Al Zahra Dist. Jeddah
	Tel. No. 02 – 6620 296 / 6620 396 / 6620 279 Ext: 5029
	Fax No. 02 – 256 2973
	Direct Line 02 - 6905129

Date : 15 / 09 / 2008	Contact Person : Dr. Alaa El Shreef - Dr. Mohy khalil
Pages : 1 : 13	E mail Address : xray1.jed@sghgroup.net - bab@sghgroup.net
Vender :	Mobile : 00966 505 640 079 - 00966 561 052 512
Attention : Sales Manager	
Enclosed please find our Tender Document for DENTAL UNIT DEPARTMENT	
Please submit your best prices for the following items complete specification, Catalogue copy & C.V for units	
Your offer in witting should reach to the undersign before 15 / Dec / 2008	
You may quote your equivalent product if you are unable to offer the specification item	

General Requirements for the Negotiable Tender

- 1- The Manufacture has to offer the Highest & Latest model in the required Item with all the Encouraging specification which let the dean of the tender to choose the Product.
- 2- The offer must contain all Hardware & Software available At time of tender.
- 3- The Manufacture will apply all the Upgrade will be added to the unit Free of Charge.
- 4- The Manufacture will offer clear C.V of the Same Model in his quotation with year of production.
- 5- Warrantee is (5) years including labor & Spare parts & Accessories needed for the perfect work of the unit.
- 6- The Manufacture will assure the wok of the unit during the Warrantee with Minimum of 320 days / Year.
- 7- It will be Service Contract for (5) years after the warrantee period in the offer of the manufacturer.
- 8- It's better for the Manufacturer to offer Basic Unit & alternative for each item.
- 9- It's well known that The Company which will win the Tender is going to Sign the Contract with . Sgh Group.
- 10- This Contract Details will be Discusses with the Company after The Evaluation with the whole Higher Management SGH Group.
- 11- The specification in the book is the nearly accepted but not restricted so each Company Can offer the best Models which Help doctors & Staff to Offer the Best Benefits for the Patient.

Dr. MOHY KHALIL
bab@sghgroup.net

INDEX

DEPARTMENT: DENTAL CLINIC

S.No	Items Descriptions	Qty	No. of pages	Manufacture	Origin	Model	Unit Price	Total Price
1	DENTAL UNIT COMPLETE	2	1					
2	CABINET SET OF 6 UNITS	2	1					
3	X-RAY WALL MOUNTED WITH MULTITIMER P	2	1					
4	X-RAY VIEWER WITH MAGIFIER	2	1					
5	APRON STAR USA 5MM	2	1					
6	DARK ROOM MANUAL PROCESSOR	2	1					
7	AMALGAMATOR.	2	1					
8	LIGHT CURE UNIT.	2	1					
9	DENTAL MOBILE UNIT PROCART II	2	1					
10	INTRA ORAL CAMERA	2	1					
11	R V G "DIGITAL VIEW FOR X- RAY"	2	1					

DENTAL UNIT COMPLETE**ITEM NO: 1**

Specification For: DENTAL UNIT COMPLETE	QTY
<p style="text-align: center;"><u>DENTAL UNIT COMPLETE</u></p> <ul style="list-style-type: none"> - 2 Titanium Power Fiber optic turbine w/4 way spray - 1 Implantor steril fiber optic micro motor with contra angle and straight head - 1 Multi steril L warm syring (Doctor Table) - 1 Multi steril cold syring (assistant table) - 1 Minitom 2 surgery with power and time programe - 1 Piezo steril electric scaller with 4 tips and endo function - 1 Foot control - Luna operating light - Pragma patient chair - Selene Lamp - 2 Vector stool (Doctor + Assistant) - Auto steril system - Automatic system for water and cup filler disinfection - Connection for telediagnosis and teleassistant. - Cardio alert - Physiological solution - Metasys centerifugal amalgam separators MSTI complete - Auto Diagnosis system. - HPA system inlet compressed air to a 220 micron filter - Tad system anti-stagnation system - Plisteril Basin. - Voltage: 110V-220V 60 Hz 	2

CABINET SET OF 6 UNITS.**ITEM NO: 2**

Specification For: CABINET SET OF 6 UNITS.	QTY
<p data-bbox="517 584 912 618" style="text-align: center;"><u>CABINET SET OF 6 UNITS.</u></p> <ul data-bbox="201 651 1257 887" style="list-style-type: none">- Rear wall paneling for free standing installation with table top type- Wash Basin element with enamel basin knee operated sliding front panel, waste basket inside the cabinet- Dentist element 2 drawers 6 cm, 1 conservative compartment 12 cm, 3 drawers 12 cm- Garage with pull out shelf(25cm wide), 2 drawers 6 cm, 2 drawers 12 cm, 1 drawer 24 cm- Element with 30 cm Universal pull-out drawer for additional equipment, 1 drawer 12 cm, 1 drawer 18 cm	2

X-RAY WALL MOUNTED WITH MULTITIMER P**ITEM NO: 3**

Specification For: X-RAY WALL MOUNTED WITH MULTITIMER P	QTY
<p data-bbox="352 584 1075 618" style="text-align: center;"><u>X-RAY WALL MOUNTED WITH MULTITIMER P</u></p> <ul data-bbox="199 651 1083 976" style="list-style-type: none">- X-ray machine, with support arm 1810 mm- X-ray tube SR 60/7L- Focal spot: 0.7- Total filters: Min. 2.5 mm Al- Exposure voltage: 60 kv multipulse- Exposure current: 7 mA- Power Requirement: 1.1 kW- Fuse: 16A, slow-blow- Exposure time: 0.01-3.2 electronically controlled, selectable digital mode- Remote control- Voltage: 110V-220V 60 Hz	2

X-RAY VIEWER WITH MAGIFIER.**ITEM NO: 4**

Specification For X-RAY VIEWER WITH GLASS MAGIFIER.	QTY
<p data-bbox="400 584 1027 618" style="text-align: center;"><u>X-RAY VIEWER WITH GLASS MAGIFIER.</u></p> <ul data-bbox="199 651 1241 712" style="list-style-type: none"><li data-bbox="199 651 1241 712">- Table wall x-ray viewer 40x20 cm, made of steel sheeting heat enameled, ideal for the illumination of standard radiograph.	2

APRON STAR USA 5MM**ITEM NO: 5**

Specification For: <u>APRON STAR USA 5MM</u>	QTY
<p data-bbox="536 584 895 618" style="text-align: center;"><u>APRON STAR USA 5MM</u></p> <ul data-bbox="199 651 1222 768" style="list-style-type: none">- X-ray apron for operator or technician. The apron perfectly covers the shoulders and body.- Lead protection: mms. 0.3- Size: 120x60 cms	2

DARK ROOM MANUAL PROCESSOR.**ITEM NO: 6**

Specification For: DARK ROOM MANUAL PROCESSOR.	QTY
<p data-bbox="437 584 991 618" style="text-align: center;"><u>DARK ROOM MANUAL PROCESSOR.</u></p> <ul data-bbox="201 651 1222 741" style="list-style-type: none">- X-ray processing developing box designed to eliminate waste of time and assure the operator a perfect and quick treatment of endodontic and surgery x-ray films- Dimension: 32x44x29 cm	2

AMAGMATOR.**ITEM NO: 7**

Specification For: AMAGMATOR.	QTY
<p data-bbox="600 551 831 584" style="text-align: center;"><u>AMAGMATOR.</u></p> <ul data-bbox="201 613 791 734" style="list-style-type: none">- Microprocessor controlled, extremely accurate.- Touch pad controls, LED display- Programmable pre-set memories- Spring tensioned, molded capsule holder	2

LIGHT CURE UNIT.**ITEM NO: 8**

Specification For: LIGHT CURE UNIT.	QTY
<p data-bbox="568 551 863 584" style="text-align: center;"><u>LIGHT CURE UNIT.</u></p> <ul data-bbox="201 618 1002 734" style="list-style-type: none">- Halogen-powered light curing unit.- Fully electronic timer in 1/10th of a minute- Digital read out with "countdown" feature audible completion alert- Voltage: 110V-220V 60 Hz	2

DENTAL MOBILE UNIT PROCART II**ITEM NO: 9**

Specification For: DENTAL MOBILE UNIT PROCART II	QTY
<p style="text-align: center;"><u>DENTAL MOBILE UNIT PROCART II</u></p> <ul style="list-style-type: none"> - Compressor: - Oil-free 1/3 hp motor - Working pressure: 70 to 100 psi - Flow rate: 2.5 cfm - Sound level: 48 dB - Air reservoir: 4 liters - Automatic air tank purge - Vacuum Motor: - Oil-free 1/3 hp motor - Adjustable vacuum: 0 to 21 " Hg - Vacuum flow: 4.6 cfm - Waste receptacle: 4 liters - Sound level: 48 dB - Automatic overflow shut off aseptic filtered exhaust large volume solids trap automatic waste removal - Water Module: - 1.5 liter reservoir - Pressurized system: 30 to 40 lbs - Aseptic non-retracting HPC coolant Hand piece flush system External quick connects for water and air. - Voltage: 110V-220V 60 Hz 	2

INTRA ORAL CAMERA

ITEM NO: 10

Specification For: INTRA ORAL CAMERA	QTY
<u>INTRA ORAL CAMERA</u>	2

R V G "DIGITAL VIEW FOR X- RAY

ITEM NO: 11

Specification For: R V G "DIGITAL VIEW FOR X- RAY	QTY
<u>R V G DIDITAL VIEW FOR X- RAY</u>	2