

	Bait Al Batterjee Medical Trading Co.
	Saudi German Hospital Group
	Al Zahra Dist. Jeddah
	Tel. No. 02 – 6620 296 / 6620 396 / 6620 279 Ext: 5029
	Fax No. 02 – 256 2973
	Direct Line 02 - 6905129

Date : 15 / 09 / 2008	Contact Person : Dr. Alaa El Shreef - Dr. Mohy khalil
Pages : 1 : 12	E mail Address : xray1.jed@sghgroup.net - bab@sghgroup.net
Vender :	Mobile : 00966 505 640 079 - 00966 561 052 512
Attention : Sales Manager	
Enclosed please find our Tender Document for DERMATOLOGY DEPARTMENT	
Please submit your best prices for the following items complete specification, Catalogue copy & C.V for units	
Your offer in witting should reach to the undersign before 15 / Dec / 2008	
You may quote your equivalent product if you are unable to offer the specification item	

General Requirements for the Negotiable Tender

- 1- The Manufacture has to offer the Highest & Latest model in the required Item with all the Encouraging specification which let the dean of the tender to choose the Product.
- 2- The offer must contain all Hardware & Software available At time of tender.
- 3- The Manufacture will apply all the Upgrade will be added to the unit Free of Charge.
- 4- The Manufacture will offer clear C.V of the Same Model in his quotation with year of production.
- 5- Warrantee is (5) years including labor & Spare parts & Accessories needed for the perfect work of the unit.
- 6- The Manufacture will assure the wok of the unit during the Warrantee with Minimum of 320 days / Year.
- 7- It will be Service Contract for (5) years after the warrantee period in the offer of the manufacturer.
- 8- It's better for the Manufacturer to offer Basic Unit & alternative for each item.
- 9- It's well known that The Company which will win the Tender is going to Sign the Contract with SGH Group.
- 10- This Contract Details will be Discusses with the Company after The Evaluation with the whole Higher Management SGH Group.
- 11- The specification in the book is the nearly accepted but not restricted so each Company Can offer the best Models which Help doctors & Staff to Offer the Best Benefits for the Patient.

Dr. MOHY KHALIL
bab@sghgroup.net

INDEX

DEPARTMENT: DERMATOLOGY

S.No	Items Descriptions	Qty	No. of pages	Manufacture	Origin	Model	Unit Price	Total Price
1	PHOTOGENICA V-STAR . FLASH LAMP EXCITED TUNABLE DYE LASER FOR THE THEATMENT OF PORT WINE BIRTHMARKS AND OTHER VASCULAR LESIONS WITH STAND. ACCESSORIESS	1	1					
2	CO3 STATE OF THE ART SKIN RESURFACING LASER WITH ACCESS.	1	1					
3	ELECTROSURGICAL UNIT WITH STAND.ACCESS.	2	1					
4	PUVA MACHINE FOR HAND WITH TROLLY.	2	1					
5	PUVA MACHINE FOR FEET WITH TROLLY.	2	1					
6	PUVA PHOTOTHERAPY COMBINATION CABIN.	1	1					
7	STUDIO CENTER 3 WITH ALL ACCESSORIES	1	1					
8	ENHANCED THERMO -KIENTIC SLECTIVITY, SOLID STATE ALEXANDRITE HAIR REMOVING LASER WITH STAND. ACCESS.	1	1					
9	EXPRESS SCANNER FOR APOGEE	1	1					
10	ZIMMER -SMART COOL, COOLING DEVICE WITH ADAPTOR	1	1					

**PHOTOGENICA V-STAR . FLASH LAMP EXCITED TUNABLE DYE
LASER FOR THE THEATMENT OF PORT WINE BIRTHMARKS AND
OTHER VASCULAR LESIONS WITH STAND. ACCESSORIESS**

ITEM NO: 1

Specification For : PHOTOGENICA V-STAR . FLASH LAMP EXCITED TUNABLE DYE LASER FOR THE THEATMENT OF PORT WINE BIRTHMARKS AND OTHER VASCULAR LESIONS WITH STAND. ACCESSORIESS	QTY
<p align="center">PHOTOGENICA V-STAR . FLASH LAMP EXCITED TUNABLE DYE LASER FOR THE THEATMENT OF PORT WINE BIRTHMARKS AND OTHER VASCULAR LESIONS WITH STAND. ACCESSORIESS</p> <ul style="list-style-type: none"> - Flash lamp excited tunable dye laser type. - 595nm Wavelength. - Solid State Aiming Beam. - Pulse Duration: 0.5 MSEC. - Air Cooled. - Spot Size: 7mm – 20 J/cm², 10mm – 10 J/cm² - Beam Delivery: Lens Coupled Optical Fiber With Hand-piece. - Dye kit: 50,000 pulses. - Voltage: 110V- 220V 60 Hz. 	1

CO3 STATE OF THE ART SKIN RESURFACING LASER WITH ACCESS.**ITEM NO: 2**

Specification For: CO3 STATE OF THE ART SKIN RESURFACING LASER WITH ACCESS	QTY
<p><u>CO3 STATE OF THE ART SKIN RESURFACING LASER WITH ACCESS.</u></p> <ul style="list-style-type: none"> - Flash lamp – State Er: YAG Laser; 2940nm wavelength. - Output mode: Multimode. - Average power: 15 watts - Pulse Duration : 0.5 msec., 7 msec., 10 msec. - Repetition rate : 2 – 10 Hz Adjustable. - Pulsing Controls: Finger switch of Footswitch with Aiming Beam including: - - 1ea 7mm Spot size Metal Tip hand piece. - 1ea 5mm Spot size metal tip hand piece. - 1ea 3mm spot size metal tip hand piece. - 1ea Footswitch of Finger switch Assembly - 1ea Patient Safety Goggles (2904nm). - 1ea Laser warning sign. - 1ea Interlock plug. - 1ea Set operating keys. - Voltage: 110V- 220V 60 Hz. 	1

ELECTROSURGICAL UNIT WITH STAND.ACCESS**ITEM NO: 3**

Specification For ELECTROSURGICAL UNIT WITH STAND.ACCESS	QTY
<p data-bbox="341 479 1086 517" style="text-align: center;"><u>ELECTROSURGICAL UNIT WITH STAND.ACCESS</u></p> <ul data-bbox="199 544 895 719" style="list-style-type: none">- Frequency: 3.8 MHz, Power Output: 140 Watts.- UL Listed, FDA 510K- Autoclavable Hand pieces, Electrodes and antenna plat.- Footswitch activated, with optional finger switch control.- Effective, User-Friendly design.- Voltage: 110V- 220V 60 Hz.	2

PUVA MACHINE FOR HAND WITH TROLLY.**ITEM NO: 4**

Specification For: PUVA MACHINE FOR HAND WITH TROLLY.	QTY
<p data-bbox="379 510 1050 544" style="text-align: center;"><u>PUVA MACHINE FOR HAND WITH TROLLY.</u></p> <ul data-bbox="199 573 1093 719" style="list-style-type: none">- 6 lamps Philips 36W>PL36W/09 UVA, 4 or 8 lamps PL9W/12 UVB.- Radiation spectrum UVA between 320 and 370 nm.- Radiation spectrum UVB between 285 &350 nm with peak at 310/315mm- 1 Ventilator to cool the lamps, Timer adjustable at the second.- Voltage: 110V- 220V 60 Hz.	2

PUVA MACHINE FOR FEET WITH TROLLY.**ITEM NO: 5**

<u>Specification For: PUVA MACHINE FOR FEET WITH TROLLY..</u>	QTY
<u>PUVA MACHINE FOR FEET WITH TROLLY.</u> <ul style="list-style-type: none">- 12 lamps Philips 36W>PL36W/09N UVA, 4 lamps PL9W/12UVB.- Radiation spectrum UVA between 320 and 370 nm.- Radiation spectrum UVB between 285 &350 nm with peak at 310/315mm- 2 Ventilator to cool the lamps, Timer adjustable at the second.- Voltage: 110V- 220V 60 Hz.	2

PUVA PHOTOTHERAPY COMBINATION CABIN.**ITEM NO: 6**

Specification For: PUVA PHOTOTHERAPY COMBINATION CABIN.	QTY
<p data-bbox="352 510 1075 544" style="text-align: center;"><u>PUVA PHOTOTHERAPY COMBINATION CABIN.</u></p> <ul data-bbox="199 573 1098 808" style="list-style-type: none">- Exterior dimension: 1240 x 1286 x 1990 mm- Interior dimension: 890 x 890 x 1870 mm- Ventilation treatment field: Integral ventilation by low-noise tangential fans- 32 lamps UV-A, type UV-A 100W, Philips or 48 lamps UV-A.- 16 lamps UV-B, type TL01, Philips- Consumption: 502 kW , Fuse: 16A/25A- Produced as Norm IEC 601-1, 380/220V 60 Hz- Voltage: 110V- 220V 60 Hz.	1

STUDIO CENTER 3 WITH ALL ACCESSORIES

ITEM NO: 7

Specification For: STUDIO CENTER 3 WITH ALL ACCESSORIES	QTY
<p data-bbox="373 510 1054 544" style="text-align: center;"><u>STUDIO CENTER 3 WITH ALL ACCESSORIES</u></p> <ul data-bbox="199 573 1086 723" style="list-style-type: none">- Deep cleansing of the skin with the regional fornicator- Evaporation with the original Vapozone- Ultra-fine spraying facility- Vacuum treatment for removing excess sebum and for vacuum massage- Voltage: 110V- 220V 60 Hz.	1

**ENHANCED THERMO –KIENTIC SLECTIVITY , SOLID STATE
ALEXANDRITE HAIR REMOVING LASER WITH STAND . ACCESS.**

ITEM NO: 8

Specification For: ENHANCED THERMO –KIENTIC SLECTIVITY , SOLID STATE ALEXANDRITE HAIR REMOVING LASER WITH STAND . ACCESS	QTY
<u>ENHANCED THERMO –KIENTIC SLECTIVITY , SOLID STATE ALEXANDRITE HAIR REMOVING LASER WITH STAND . ACCESS.</u>	1

ITEM NO: 9 **EXPRESS SCANNER FOR APOGEE**

Specification For: EXPRESS SCANNER FOR APOGEE	QTY
<u>EXPRESS SCANNER FOR APOGEE</u>	1

ZIMMER –SMART COOL, COOLING DEVICE WITH ADAPTOR

ITEM NO: 10

Specification For: ZIMMER –SMART COOL, COOLING DEVICE WITH ADAPTOR	QTY
<u>ZIMMER –SMART COOL, COOLING DEVICE WITH ADAPTOR</u>	1