

	Bait Al Batterjee Medical Trading Co.
	Saudi German Hospital Group
	Al Zahra Dist. Jeddah
	Tel. No. 02 – 6620 296 / 6620 396 / 6620 279 Ext: 5029
	Fax No. 02 – 256 2973
	Direct Line 02 - 6905129

Date : 15 / 09 / 2008	Contact Person : Dr. Alaa El Shreef - Dr. Mohy khalil
Pages : 1 : 25	E mail Address : xrav1.jed@sghgroup.net - bab@sghgroup.net
Vender :	Mobile : 00966 505 640 079 - 00966 561 052 512
Attention : Sales Manager	
Enclosed please find our Tender Document for ENDOSCOPY DEPARTMENT	
Please submit your best prices for the following items complete specification, Catalogue copy & C.V for units	
Your offer in witting should reach to the undersign before 15 / Dec / 2008	
You may quote your equivalent product if you are unable to offer the specification item	

General Requirements for the Negotiable Tender

- 1- The Manufacture has to offer the Highest & Latest model in the required Item with all the Encouraging specification which let the dean of the tender to choose the Product.
- 2- The offer must contain all Hardware & Software available At time of tender.
- 3- The Manufacture will apply all the Upgrade will be added to the unit Free of Charge.
- 4- The Manufacture will offer clear C.V of the Same Model in his quotation with year of production.
- 5- Warrantee is (5) years including labor & Spare parts & Accessories needed for the perfect work of the unit.
- 6- The Manufacture will assure the wok of the unit during the Warrantee with Minimum of 320 days / Year.
- 7- It will be Service Contract for (5) years after the warrantee period in the offer of the manufacturer.
- 8- It's better for the Manufacturer to offer Basic Unit & alternative for each item.
- 9- It's well known that The Company which will win the Tender is going to Sign the Contract with SGH Group.
- 10- This Contract Details will be Discusses with the Company after The Evaluation with the whole Higher Management SGH Group.
- 11- The specification in the book is the nearly accepted but not restricted so each Company Can offer the best Models which Help doctors & Staff to Offer the Best Benefits for the Patient.

Dr. MOHY KHALIL
bab@sghgroup.net

INDEX

DEPARTMENT: ENDOSCOPY

S.No	Items Descriptions	Qty	No. of pages	Manufacture	Origin	Model	Unit price	Total price
1	PENDANT CEILING SINGLE ARM.	1	1					
2	MOBILE OPERATING TABLE WITH ACCESSORIES.	1	1					
3	GASTROSCOPE SYSTEM AND COMPLETE ACCESS.	1	1					
4	DUODENOSCOPE	1	1					
5	COLONOSCOPE	1	1					
6	GASTROSCOPE FIBERSCOPE	1	1					
7	VIDEO PROCESSOR WITH MONITOR CABLE	1	1					
8	Light Source With Water Container	1	1					
9	Video Monitor Sony #PMV-20M2MDE	1	1					
10	Diathermy Unit #PSD-20	1	1					
11	Color printer Sony	1	1					
12	Suction Pump #KV-5	1	1					
13	Digital Recorder	1	1					
14	Disinfector Manual.	1	1					
15	Mobile Workstation	1	1					
16	Washer	1	1					
17	Maintenance Unit	1	1					
18	Heat Probe Unit	1	1					
19	Video Converter.	1	1					
20	Double Accessory Hanger Unit	1	1					
21	HIGH ENDOSCOPY CUPBOARD W/FASTEMER	1	1					
22	HIGH STORAGE CUPBOARD W/LOCK	1	1					
23	PULSE OXYMETER	1	1					

PENDANT CEILING SINGLE ARM.**ITEM NO: 1**

Specification For: PENDANT CEILING SINGLE ARM.	QTY
<p style="text-align: center;"><u>PENDANT CEILING SINGLE ARM.</u></p> <ul style="list-style-type: none"> - Parking and working position are fundamentally different - Speedier change-over of medial equipment and instruments - Quick, easy swiveling in case of emergency - Supply of high and low tension current as well as medical gases to the surgical instruments. - Clearance height of 2 m - Modular equipping of appliances of differing weights and sizes - Variable work positions when patients are brought in and out. - The basic components consist of: Box, Shelves, Drawers, Keyboard/writing drawers, Tilt/Swivel section, I.V Poles, docking/lifting system for apparatus, Electrical connections, Medical gas connection, special apparatus, communication and data lines. - Air motor - Voltage: 110V- 220V 60 Hz 	1

MOBILE OPERATING TABLE WITH ACCESSORIES.**ITEM NO: 2**

Specification For: MOBILE OPERATING TABLE WITH ACCESSORIES.	QTY
<p style="text-align: center;"><u>MOBILE OPERATING TABLE WITH ACCESSORIES.</u></p> <ul style="list-style-type: none"> - Mobile operating table with Pedal-Operated Hydraulic Height adjustments including radiotranslucent/electrically, conductive SFC padding, fixed with Velcro straps. Trendelenburg/Reverse Trendelenburg position and back plate adjustment by manual pressure supported adjustments. Lateral tilt adjustment (left/right) from the receptacles at the table frame. - Table top frame column casing made entirely of CrNi steel. Solid base construction. Base covers made of special heavy-duty plastic material, resistant to impact, breakage and disinfectants. - The operating table is mobile on 4 large castors of 100 mm diameter and can be locked by means of double swivel castors at the head end by using a central pedal. - Length without head rest: 1745 mm - Length with head rest: 2025 mm - Width with side rails: 605 mm - Width without side rails: 540 mm - Height, including padding: 780 to 1030 mm - Trendelenburg and reverse Trendelenburg: 30° - Lateral tilt left /right: 20° - Back plate upward/downward: 70°/40° - Leg plate downward: 25/90° - Weight: 156 kg - Voltage: 110V- 220V 60 Hz 	1

GASTROSCOPE SYSTEM AND COMPLETE ACCESS**ITEM NO: 3**

Specification For: GASTROSCOPE SYSTEM AND COMPLETE ACCESS	QTY
<p style="text-align: center;"><u>GASTROSCOPE SYSTEM AND COMPLETE ACCESS</u></p> <ul style="list-style-type: none"> - Consists of: 1 ea Main Channel Conn Assy Type 2, 1 ea Water Container - Optical system: <ul style="list-style-type: none"> - Field of view: 140° - Direction of view: 0° forward viewing - Depth of field: 3 to 100 mm - Distal End: <ul style="list-style-type: none"> - Outer diameter: 9.0 mm - Insertion tube: <ul style="list-style-type: none"> - Outer diameter: 9.0 mm - Bending section: <ul style="list-style-type: none"> - Range of distal end: Up 210°, Down 90° - Bending: Right and left, each 100° - Working Length: 1,030 mm - Total Length: 1,335 mm - Inter diameter: 2.8 mm - Minimum visible distance: 3 mm - Voltage: 110V-220V 60 Hz 	1

DUODENOSCOPE.**ITEM NO: 4**

Specification For: DUODENOSCOPE	QTY
<p style="text-align: center;"><u>DUODENOSCOPE.</u></p> <ul style="list-style-type: none"> - Consists of: 1 ea Main Channel Conn Assy Type 2, 1 ea Water Container and 1 ea Encapsulated Card. - Optical system: <ul style="list-style-type: none"> - Field of view: 100° - Direction of view: 5° backward-oblique - Depth of field: 5 to 60 mm - Distal End: <ul style="list-style-type: none"> - Outer diameter: 13.5 mm - Insertion tube: <ul style="list-style-type: none"> - Outer diameter: 12 mm - Bending section: <ul style="list-style-type: none"> - Range of distal end: Up 210°, Down 90° - Bending: Right 110o and left 90° - Working Length: 1,240 mm - Total Length: 1,540 mm - Inter diameter: 4.2 mm - Minimum visible distance: 10 mm - Voltage: 110V-220V 60 Hz 	1

COLONOSCOPE**ITEM NO: 5**

Specification For: COLONOSCOPE	QTY
<p style="text-align: center;"><u>COLONOSCOPE.</u></p> <ul style="list-style-type: none"> - Consists of: 1 ea Main Channel Conn Assy Type 2, 1 ea Water Container. - Optical system: <ul style="list-style-type: none"> - Field of view: 140° - Direction of view: 0° forward viewing - Depth of field: 5 to 100 mm - Distal End: <ul style="list-style-type: none"> - Outer diameter: 13.2 mm - Insertion tube: <ul style="list-style-type: none"> - Outer diameter: 12.9 mm - Bending section: <ul style="list-style-type: none"> - Range of distal end: Up and Down each 180° - Bending: Right and left 160° - Working Length: L: 1,680 mm, I: 1,330 mm - Total Length: L: 1,995 mm, I: 1,645 mm <ul style="list-style-type: none"> - Inter diameter: 3.7 mm - Minimum visible distance: 5 mm - Voltage: 110V-220V 60 Hz 	1

GASTROSCOPE FIBERSCOPE.**ITEM NO: 6**

Specification For: GASTROSCOPE FIBERSCOPE.	QTY
<p style="text-align: center;"><u>GASTROSCOPE FIBERSCOPE.</u></p> <ul style="list-style-type: none"> - Optical system: - Field of view: 120° - Direction of view: 0° forward viewing - Depth of field: 3 to 100 mm - Distal End: - Outer diameter: 11.2 mm - Insertion tube: - Outer diameter: 11.3 mm - Bending section: - Range of distal end: Up 210°, Down 90° - Bending: Right 100° and left 100° - Working Length: 1,025 mm - Total Length: 1,345 mm - Inter diameter: 3.7 mm - Biopsy Forceps: Minimum visible distance: 3 mm from distal end - Voltage: 110V-220V 60 Hz 	1

VIDEO PROCESSOR WITH MONITOR CABLE

ITEM NO: 7

Specification For: Video Processor With Monitor Cable.	QTY
<u>VIDEO PROCESSOR WITH MONITOR CABLE</u>	1

Light Source With Water Container.

ITEM NO: 8

Specification For: Light Source With Water Container.	QTY
<u>Light Source With Water Container.</u>	1

Video Monitor Sony #PMV-20M2MDE

ITEM NO: 9

Specification For: Video Monitor Sony #PMV-20M2MDE	QTY
<u>Video Monitor Sony #PMV-20M2MDE</u>	1

Diathermy Unit #PSD-20**ITEM NO: 10**

Specification For: Diathermy Unit #PSD-20	QTY
<p data-bbox="544 562 884 600" style="text-align: center;"><u>Diathermy Unit #PSD-20</u></p> <ul data-bbox="199 658 395 748" style="list-style-type: none">- Patient plate- Adaptor- P-Adaptor	1

Color printer Sony

ITEM NO: 11

Specification For: Color printer Sony	QTY
<p style="text-align: center;"><u>Color printer Sony</u></p> <p>- Large size print Sony</p>	1

Suction Pump #KV-5**ITEM NO: 12**

Specification For: Suction Pump #KV-5	QTY
<p data-bbox="571 555 858 591" style="text-align: center;"><u>Suction Pump #KV-5</u></p> <ul data-bbox="199 651 529 831" style="list-style-type: none">- Suction Jar Universal 2L- Lid for use without Liner- Tubing 2M- Tube to Filter 900 mm- Jar Mounting Kit- Filter Keyvac	1

Digital Recorder.**ITEM NO: 13**

Specification For: Digital Recorder.	QTY
<p data-bbox="598 521 831 560" style="text-align: center;"><u>Digital Recorder.</u></p> <ul data-bbox="199 616 486 739" style="list-style-type: none">- Remote Control Unit- Video Lead 1.5 m- Video Cable super- Tape 64 Mins.	1

Disinfector Manual.

ITEM NO: 14

Specification For: Disinfector Manual.	QTY
<u>Disinfector Manual.</u>	1

Mobile Workstation.

ITEM NO: 15

Specification For: Mobile Workstation.	QTY
<p><u>Mobile Workstation.</u></p> <p>- Monitor Tray</p>	1

Washer.**ITEM NO: 16**

Specification For: Washer.	QTY
<p data-bbox="657 521 770 560" style="text-align: center;"><u>Washer.</u></p> <ul data-bbox="199 589 528 734" style="list-style-type: none">- Air Filter- Water Filter- Gas Filter- Channel Cleaning tubes- Leak test Air tube	1

Maintenance Unit.**ITEM NO: 17**

Specification For Maintenance Unit.	QTY
- Leakage Tester <u>Maintenance Unit.</u>	1

Heat Probe Unit.**ITEM NO: 18**

Specification For Heat Probe Unit.	QTY
<u>Heat Probe Unit.</u> <ul style="list-style-type: none">- Bottle for HPU- Pump for HPU	1

Video Converter.

ITEM NO: 19

Specification For Video Converter.	QTY
<u>Video Converter.</u>	1

Double Accessory Hanger Unit.

ITEM NO: 20

Specification For Double Accessory Hanger Unit.	QTY
<u>Double Accessory Hanger Unit.</u>	1

HIGH ENDOSCOPY CUPBOARD W/FASTEMER.**ITEM NO: 21**

Specification For HIGH ENDOSCOPY CUPBOARD W/FASTEMER.	QTY
<u>HIGH ENDOSCOPY CUPBOARD W/FASTEMER.</u> - 2 hinged doors, right door with bowed bar handle, 2 extensible partition wall 6 special hooks for suspending fiber endoscopes.	1

HIGH STORAGE CUPBOARD W/LOCK.**ITEM NO: 22**

Specification For: HIGH STORAGE CUPBOARD W/LOCK.	QTY
<u>HIGH STORAGE CUPBOARD W/LOCK.</u> <ul style="list-style-type: none">- 2 hinged doors with bow-shaped handle and magnetic lock, 6 adjustable shelves cupboard made of double walled power coated S/S	1

PULSE OXYMETER

ITEM NO: 23

Specification For: PULSE OXYMETER.	QTY
<u>PULSE OXYMETER</u>	1