

	Bait Al Batterjee Medical Trading Co.
	Saudi German Hospital Group
	Al Zahra Dist. Jeddah
	Tel. No. 02 – 6620 296 / 6620 396 / 6620 279 Ext: 5029
	Fax No. 02 – 256 2973
	Direct Line 02 - 6905129

Date : 15 / 09 / 2008	Contact Person : Dr. Alaa El Shreef - Dr. Mohy khalil
Pages : 1 : 6	E mail Address : xray1.jed@sghgroup.net - bab@sghgroup.net
Vender :	Mobile : 00966 505 640 079 - 00966 561 052 512
Attention : Sales Manager	
Enclosed please find our Tender Document for OB.GYNIC DEPARTMENT	
Please submit your best prices for the following items complete specification, Catalogue copy & C.V for units	
Your offer in witting should reach to the undersign before 15 / Dec / 2008	
You may quote your equivalent product if you are unable to offer the specification item	

General Requirements for the Negotiable Tender

- 1- The Manufacture has to offer the Highest & Latest model in the required Item with all the Encouraging specification which let the dean of the tender to choose the Product.
- 2- The offer must contain all Hardware & Software available At time of tender.
- 3- The Manufacture will apply all the Upgrade will be added to the unit Free of Charge.
- 4- The Manufacture will offer clear C.V of the Same Model in his quotation with year of production.
- 5- Warrantee is (5) years including labor & Spare parts & Accessories needed for the perfect work of the unit.
- 6- The Manufacture will assure the wok of the unit during the Warrantee with Minimum of 320 days / Year.
- 7- It will be Service Contract for (5) years after the warrantee period in the offer of the manufacturer.
- 8- It's better for the Manufacturer to offer Basic Unit & alternative for each item.
- 9- It's well known that The Company which will win the Tender is going to Sign the Contract with SGH Group.
- 10- This Contract Details will be Discusses with the Company after The Evaluation with the whole Higher Management SGH Group.
- 11- The specification in the book is the nearly accepted but not restricted so each Company Can offer the best Models which Help doctors & Staff to Offer the Best Benefits for the Patient.

Dr. MOHY KHALIL
bab@sghgroup.net

INDEX

DEPARTMENT: OB GYNIC

S. No	Items Descriptions	Qty	No. of pages	Manufacture	Origin	Model	Unit price	Total Price
1	GYNAECOLOGY EXAMINATION TABLE WITH STANDARD ACCESS..	3	1					
2	VASCULAR DOPPLER	1	1					
3	ULTRASOUND WITH STAND . ACCESS.W/ENDO-CAVITY PROBE	4	1					
4	AUTOMATIC VIDEO COLPOSCOPE WITH STANDARD ACCESS.	1	1					

GYNAECOLOGY EXAMINATION TABLE WITH STANDARD ACCESS**ITEM NO: 1**

Specification For: GYNAECOLOGY EXAMINATION TABLE WITH STANDARD ACCESS	QTY
<p><u>GYNAECOLOGY EXAMINATION TABLE WITH STANDARD ACCESS.</u></p> <ul style="list-style-type: none"> - Electrically powered Trendelenburg position and in addition, electrically powered adjustment of seat section, height adjustment also electrically powered from 680-900 mm, mounting height: 630 mm. - Seat, back rest and head rest in moulded upholstery; stainless steel side rails 25x10 mm fitted to the seat section stainless steel - Rinsing basin, paper roll holder, remote control facility by means of foot-operated switch. - Base plate with 4 leveling screws, plastic claddings and covers , telescopic steel casing and seat frame plastic enamel coated. - Weight: 165 kg - Voltage: 110V-220V 60 Hz 	3

VASCULAR DOPPLER**ITEM NO: 2**

Specification For: VASCULAR DOPPLER	QTY
<p data-bbox="560 555 895 591" style="text-align: center;"><u>VASCULAR DOPPLER</u></p> <ul data-bbox="177 622 595 797" style="list-style-type: none">- Hand-Held Design- Multi frequency: 2,4,5,8,10 MHz- Probe button: Power ON/OFF- Battery life: 3 hours- Power: DC 9Volt- Dimension: 75x140x25 mm	1

ULTRASOUND WITH STAND . ACCESS.W/ENDO-CAVITY PROBE**ITEM NO: 3**

Specification For: ULTRASOUND WITH STAND . ACCESS.W/ENDO-CAVITY PROBE	QTY
<p><u>ULTRASOUND WITH STAND . ACCESS.W/ENDO-CAVITY PROBE</u></p> <ul style="list-style-type: none"> - Highest resolution - Multitude of uses with an imaging bandwidth of 1 to 15 MHz - Doppler system ensures highly sensitive Doppler, CFM and Power Doppler images, CW is supported on imaging probes. - True Echo Harmonics feature particularly useful for difficult to scan patients. - Broad rang of task designed linear, convex and unique 360o rotating transducers for many special applications including image-guided biopsies. - Full 14" flicker-free monitor. - Ergonomic and compact design with height adjustable keyboard and monitor and logically grouped controls. - Integrated 3D - Palm Control Unit - Simultaneous split-screen feature display two active images on the monitor at the same time. - Multifrequency (2.7, 3.5 & 5 MHz) Convex Array Transducer for general Abdominal and Pelvic Scanning. - Multifrequency (5,6.5 & 7.5 MHz) Transvaginal Transducer with 160° scan angle - Voltage: 110V-220V 60 Hz 	4

AUTOMATIC VIDEO COLPOSCOPE WITH STANDARD ACCESS**ITEM NO: 4**

Specification For: AUTOMATIC VIDEO COLPOSCOPE WITH STANDARD ACCESS.	QTY
<p data-bbox="252 600 1198 633" style="text-align: center;"><u>AUTOMATIC VIDEO COLPOSCOPE WITH STANDARD ACCESS</u></p> <ul data-bbox="177 667 863 813" style="list-style-type: none">- Basic Colposcope with Vertical stand RS232 connector- Focal Length: 30 mm- Magnification: 4.5 x-25x- Field of View:66mm-14mm- Depth of field: 112mm-5mm <p data-bbox="129 813 496 846">- <u>Voltage: 110V-220V 60 Hz</u></p>	1